


DETAIL B
SCALE 10 : 1

1. ALL CIRCLED NOTES, DIMENSIONS, AND FEATURES ARE TO BE INSPECTED.
2. PRODUCT SHALL BE FREE OF GREASE, OILS, DIRTS, OR GELS.
3. ALL VISUAL INSPECTIONS SHALL BE PERFORMED WITH NORMAL OR CORRECTED TO NORMAL VISION UNDER NORMAL ROOM LIGHTING AT A DISTANCE OF 12-18 INCHES.
4. INSPECTION SHALL OCCUR PER AQL 4.0, C=0.

UNLESS OTHERWISE SPECIFIED:		<small>PROPRIETARY AND CONFIDENTIAL</small>  <small>THE INFORMATION CONTAINED IN THIS DRAWING IS THE PROPERTY OF MIDWEST INTERVENTIONAL SYSTEMS INC. ANY USE OR REPRODUCTION IN PART OR AS A WHOLE WITHOUT THE WRITTEN PERMISSION OF MIDWEST INTERVENTIONAL SYSTEMS INC IS PROHIBITED.</small>	
DIMENSIONS ARE IN INCHES TOLERANCES (mm): TOLERANCES (in): X±5 X±0.2 X.X±2.5 X.X±0.1 X.XX±0.25 X.XX±0.01 X.XXX±0.125 X.XXX±0.005 Angular: ± 1.5°		TITLE:	
MATERIAL LDPE Dow 640		Tipped Dilator	
FINISH N/A	DO NOT SCALE DRAWING	SIZE A	DOC. NO. SA0280
		UoM EA	REV 01
		EFFECTIVE DATE:	SCALE: 2:1 SHEET 2 OF 2